



Food and Drug Administration
10903 New Hampshire Avenue
Document Control Center – WO66-G609
Silver Spring, MD 20993-0002

November 2, 2016

Encore Medical, L.P.
Ms. Teffany Hutto
Manager, Regulatory Affairs
9800 Metric Blvd
Austin, Texas 78758

Re: K111629

Trade/Device Name: Reverse Humeral Socket Shell
Regulation Number: 21 CFR 888.3660
Regulation Name: Shoulder joint metal/polymer semi-constrained cemented prosthesis
Regulatory Class: Class II
Product Code: PHX, KWS
Dated: September 8, 2011
Received: September 9, 2011

Dear Ms. Hutto:

This letter corrects our substantially equivalent letter of September 16, 2011.

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical

device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to

<http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,

Lori A. Wiggins -S

for

Mark N. Melkerson

Director

Division of Orthopedic Devices

Office of Device Evaluation

Center for Devices and

Radiological Health

Enclosure

K111629

510(k) Summary

SEP 16 2011

Date: September 8, 2011**Contact Person:****Manufacturer:**

DJO Surgical (legally Encore Medical, L.P.)
 9800 Metric Blvd
 Austin, TX 78758

Teffany Hutto

Manager, Regulatory Affairs

Phone: (512) 834-6255

Fax: (512) 834-6313

Email: teffany.hutto@djosurgical.com

Product	Classification	Product Code
Turon to RSP Conversion Shell	Class II	KWS

Product Code	Regulation and Classification Name
KWS	Shoulder joint metal/polymer semi-constrained prosthesis per 21 CFR 888.3660

Description: In cases of revision surgeries to a well fixed Turon Humeral Stem, a humeral socket shell can be mated with the Turon stem to convert to a reverse shoulder application.

There is no change to the intended use or fundamental scientific technology. This includes no changes to currently cleared devices, packaging or sterilization.

Indications for Use:

The Turon to RSP Conversion Shell is indicated for revision surgeries in patients with a grossly rotator cuff deficient shoulder joint with severe arthropathy or a previously failed joint replacement with a grossly rotator cuff deficient shoulder joint. The patient's joint must be anatomically and structurally suited to receive the selected implant(s), and a functional deltoid muscle is necessary to use the device. The socket shell is only indicated for use with a well fixed Turon Humeral Stem.

Predicate Device:

- DJO Surgical Reverse® Shoulder Prosthesis - K041066
- Turon Shoulder System - K080402

Comparable Features to Predicate Device(s): Features comparable to predicate devices include the same design features, indications, sterilization, and intended use.

Non-Clinical Testing:

- Fatigue Testing was performed on the socket shell assembled with the Turon stem to demonstrate that the component would withstand minimum load conditions per ASTM standards. The results of the testing determined that the socket shell could withstand the load conditions showing equivalence to the cleared Turon socket and stem interface.
- Because the locking feature geometry of the Reverse Humeral Socket Shell is the same as the predicate RSP socket shells, Torsional Strength, Lever out Strength, and Push Out Strength performed against the RSP socket shell can be applied to the Reverse Humeral Socket Shell.
- Because the mating geometry of the morse taper connection is the same as the predicate Turon Humeral Neck, Pull-out Testing conducted with the cleared Turon stem can be applied.

The above testing demonstrated the device's ability to perform under expected conditions and is equivalent to the predicate devices.

Clinical Testing: None provided.

510(k) Number (if known): K111629

Device Name: Reverse Humeral Socket Shell

Indications for Use:

**Reverse Humeral Socket Shell
Indications for Use**

The Reverse Humeral Socket Shell is indicated for revision surgeries in patients with a grossly rotator cuff deficient shoulder joint with severe arthropathy or a previously failed joint replacement with a grossly rotator cuff deficient shoulder joint. The patient's joint must be anatomically and structurally suited to receive the selected implant(s), and a functional deltoid muscle is necessary to use the device. The socket shell is only indicated for use with a well fixed Turon Humeral Stem.


Prescription Use X
(Part 21 CFR 801 Subpart D)

AND/OR

Over-The-Counter Use _____
(21 CFR 801 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)


(Division Sign-Off)
Division of Surgical, Orthopedic,
and Restorative Devices

510(k) Number K111629